




Total Laboratory Services Ltd

Inspection Certificate

Number: 03-11-2023

Distance Plates V95005

Thickness measurement	Nominal	Average	Actual	
			Xmin	Xmax
Distance plate 0.6mm	0.6 ±0.05	0.62	0.61	0.62
Distance plate 1.0mm	1.0 ±0.05	1.00	1.00	1.04
Distance plate 1.5mm	1.5 ±0.05	1.50	1.49	1.52



* Certificate shows average for the plates in each category.


Material: Stainless Steel

Calibration Equipment:
Mitutoyo 150 mm Digital Calliper. Serial Number: 019160

Calibration Date: 20 March 2023 Re-Calibration Due: 20 March 2024

Certificate Number: 1792986

Date of Inspection: 03 November 2023
Blandford Forum, Dorset



Certified by: *P. Netherway*
Paul Netherway, Managing Director

Total Laboratory Services Ltd
14 C, Sunrise Business Park
Blandford Forum, Dorset DT11 8ST