



Total Laboratory Services Ltd

Inspection Certificate

Dynamic calibration system, type AutoCal 2

Instrument No: 1083572950

Sensor No: 1006479

Inspection Date: 21-12-2023

Inspector: P. Netherway


Measurement:	Nominal	Value	Result
Check 1:	10.0 kg \pm 0.1kg	10.0kg	Passed
Check 2:	20.0 kg \pm 0.1kg	20.0kg	Passed
Check 3:	30.0 kg \pm 0.1kg	30.0kg	Passed

Used measuring equipment (traceable to national standards):

Weight 20KG W2404

Weight 10KG W2405

Blandford Forum, Dorset



Certified by: *P. Netherway*
Paul Netherway, Managing Director

Total Laboratory Services Ltd
Unit 14 C, Sunrise Business Park
Blandford Forum, Dorset DT11 8ST