

# CERTIFICATE OF CALIBRATION

Issued by: **Compli Group Ltd.**

Unit 1, Cabot Business Village, Holyrood Close, Poole, Dorset, BH17 7BA  
www.compli.uk

Telephone: 01202 656165  
Email: sales@compli.uk

Date of Issue: 22 December 2023

Certificate Number: 385758

BS:EN ISO9001:2015

**L R Q A**

**REGISTERED COMPANY**

**LRQ 00005310**

Customer:

Total Laboratory Services (TLS)  
Unit 14C  
Sunrise Business Park  
Blandford Forum  
Dorset  
DT11 8ST

Page 1 of 1 Page  
Approved Signatory  
M. Hill



**Equipment:** Feeler Gauge Set, Metric,

**Serial No:** 2693856/324

**Conditions of Calibration:** <70 %rh & 20 °C ±2 °C

**Date of Calibration:** 22 December 2023 **Procedure No.** Pro 0065

**Basis of Calibration:** Manufacturer's Specification

**Method of Calibration:** Calibrated using a horizontal metroscope.

**Results:** As found unless specified

Feeler indication mm	Measured deviation 1=0.001 mm	Tolerance 1=0.001 mm	Result
0.05	+1.0	+/- 5	Pass
0.10	+1.4	+/- 5	Pass
0.15	+1.8	+/- 5	Pass
0.20	+1.6	+/- 5	Pass
0.25	+4.6	+/- 5	Pass
0.30	-1.0	+/- 8	Pass
0.40	+0.6	+/- 8	Pass
0.50	+2.4	+/- 8	Pass
0.60	+4.8	+/- 8	Pass
0.70	+5.6	+/- 10	Pass
0.80	+10.0	+/- 10	Pass
0.90	+7.0	+/- 10	Pass
1.00	+9.2	+/- 10	Pass

Measurement Uncertainty: +/- 0.002 mm  
Measured results are traceable to national standards via equipment  
PICL077, Certificate: 327251 ( Cal due: 05/05/2024 )

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%