




Total Laboratory Services Ltd

# Inspection Certificate

## Distance Plates Z95013

Number: 31-01-2024

Thickness measurement	Nominal	Average	Actual	
			Xmin	Xmax
Distance plate 0.6mm	0.6 ±0.05	0.58	0.56	0.59
Distance plate 1.0mm	1.0 ±0.05	1.00	0.98	1.02
Distance plate 1.5mm	1.5 ±0.05	1.50	1.49	1.51



\* Certificate shows average for the plates in each category.


Material: Stainless Steel

Calibration Equipment:  
Mitutoyo 150 mm Digital Calliper. Serial Number: 019160

Calibration Date: 20 March 2023 Re-Calibration Due: 20 March 2024

Certificate Number: 1792986

Date of Inspection: 31 January 2024  
Blandford Forum, Dorset



**Certified by:** *P. Netherway*  
Paul Netherway, Managing Director

Total Laboratory Services Ltd  
14 C, Sunrise Business Park  
Blandford Forum, Dorset DT11 8ST