

MNT



**RS Calibration**  
Calibration and Repair Service  
Serial No: 1347543/138  
Cert No: 1854175  
Cal Date: 13 Mar 2024  
Recal Due:

DPN 175 Lammas Road, Corby, Northants, NN17 9RS

# **\*\*Calibration Certificate\*\***

## **Do Not Destroy**

Calibration Certificate Attached: 1854175  
OD ref: 1232364174

Digital Stopwatch

first

## **IMPORTANT INFORMATION**

Simply detach the label in the top right hand corner of the new front sheet and apply to your instrument as required.



**RS Calibration**  
Calibration and Repair Service  
Serial No: 2A345G02  
Cert No: 123456  
Cal Date: 01 May 2014  
Recal Due:

DPN 175 Lammas Road, Corby, Northants, NN17 9RS

**For Re-Calibration of your unit please email:**  
[calibration.uk@rs-components.com](mailto:calibration.uk@rs-components.com)  
or call us on 01536 405545 to arrange free collection. Please quote serial number when returning.



# CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 13 Mar 2024

Certificate No.

1854175



Calibration and Repair Service

DPN 175, Lammas Rd,  
Weldon Industrial Est  
Corby, Northants, NN17 9RS

Tel: 01536 405545

Fax: 01536 401590

Page 1 of 2 Pages

Martyn Tiney

## Client

TOTAL LABORATORY SERVICES LTD  
BLANDFORD FORUM  
DORSET  
DT11 8ST

## Instrument

Digital Stopwatch

## Serial No.

1347543/138

## Client Reference

N/A

## Procedure ID.

811\_1814\_STOPWATCH Rev. P3

## Date of Calibration

13 Mar 2024

## Performance Status

Reported Values

## Equipment Used to Carry Out Calibration

Frequency Generator

Stop Watch Test Box

## Equipment ID.

Cal 1284

Cal 1235

The measurements reported in this certificate were carried out using equipment whose values are traceable to national standards.

The management controls of the RS Calibration Laboratory are registered under the British Standard BS EN ISO 9001 : 2015 No. RS 00362.

## Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

This certificate reports recorded values for the instrument 'As Received'.

The following calibration results relate only to the items defined above.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION



Certificate No.

1854175

Page 2 of 2 Pages

## Environment

The ambient temperature and relative humidity throughout the calibration were  $(20 \pm 2) ^\circ\text{C}$  and  $(40 \pm 20) \%RH$  respectively.

## Method

The instrument was calibrated by comparison with a GPS frequency standard.

Prior to the calibration the instrument was allowed to stabilise in the laboratory for a period of not less than 30 minutes.

## Time Interval Accuracy

Indicated Interval	UUT Displayed Value	Equivalent Value
9.89 s	00 m 09 s	9.88 s
30.02 s	00 m 30 s	30.00 s
60.02 s	01 m 00 s	60.03 s
900.01 s	15 m 00 s	900.00 s
1799.89 s	29 m 59 s	1799.90 s
3539.97 s	59 m 00 s	3540 s

Measurement uncertainties of the recorded values:

$\pm 0.10s + 1 \text{ L.S.D}$  of the display resolution

**CALIBRATED BY:- MNT**

## Reported values

The uncertainties quoted refer to the recorded values, which include any identified contribution of the instrument under test and not to the ability of the instrument to maintain it's calibration.

END OF CALIBRATION