

Rototherm Group Kenfig Industrial Estate Margam, Port Talbot, SA13 2PW, United Kingdom Tel: +44 (0) 1656 740 551 Registered in Wales No. 2570730 W: www.digitron.com	TOTAL LABORATORY SERVICES LTD UNIT 14C SUNRISE BUSINESS PARK BLANDFORD FORUM DORSET DT11 8ST UNITED KINGDOM	Customer Address:
--	--	-------------------------------

Date of Issue	28 May 2024	Certificate Number	44308
---------------	-------------	--------------------	-------

Customer Details	
Company:	TOTAL LABORATORY SERVICES LTD
Customer PO Number:	2024097
Customer/ Works Order No.:	HM65374

Instrument Details			
Product Code:	2025P7	Description:	2000 Series Absolute Manometer
Serial Number:	581078958	Range:	0 - 2 Bar
Customer ID No. (if req'd):	N/A	System Accuracy:	20°C to 50°C 0.1% rdg +0.1%fs +1 digit, -10°C to 20°C 0.15% rdg +0.15% fs +1 digit
Condition of UUT:	GOOD		

Traceability

Measurements were taken in these conditions:

Temperature (°C)	21.5
Atmospheric Pressure (Mbar)	1000

Measurements were taken using listed test equipment in table below:	UKAS Certificate No.
Test Equipment Description	Test Equipment ID
MENSOR CPC 6000	EQ.3084
	21907, 16010 to 16014-1 & 16173,16172.

Test Results

Ref mbar/bar	UUT Rdg mbar/bar	PASS/FAIL	Allowable Tolerance mbar/bar
30.0	29.8	PASS	2.1
190.0	189.9	PASS	2.3
1000	1002	PASS	4.0
1950	1949	PASS	5.0

The calibration was conducted in a temperature controlled environment maintained at 21 ± 2°C in accordance with the applicable calibration procedures of Rototherm Group using UKAS certified test equipment

Tested By:	T Gough	Calibration Date	28 May 2024
Signature:	<i>T Gough</i>	Position Held	Calibration Technician
Checked by Competent Person Supervisor, Line Manger, QA, etc.		T Gough	