

Weights Limited

Unit A5, Greenway Business Park
Winslow Road, Great Horwood
Buckinghamshire, MK17 0NP

Tel: 01296 712171
sales@weights.uk.com
www.weights.uk.com

Certificate number **47367**
29 April 2024

Total Laboratory Services Limited
Unit 14C
Sunrise Business Park
Blandford Forum
Dorset
DT11 8ST

CERTIFICATE of CALIBRATION

We hereby certify that the four stainless steel weights 50g – 5g listed below have been calibrated to fall within O.I.M.L. class F₁ tolerance.

Nominal Value	Measured Value g	Uncertainty +/- mg on UKAS Weights
50 g W11939	50.000 1	0.03
20 g W11940	20.000 12	0.025
10 g W11941	10.000 07	0.020
5 g W11942	5.000 01	0.016

The measured values reported in this certificate were determined by comparison weighing methods against our laboratory's reference standards with a hypothetical density of 8000 kg/m³ which in air of density 1.2kg/m³ would balance the nominal weight.

WEIGHTS standard class E₂ weight set number 988 certified on UKAS certificate number UN0627


Date of issue 28th June 2023 by Norfolk Calibration Services. UKAS calibration number 0260.

Traceability to National Standards is established by comparison to Norfolk Calibration Services class E₂ weight sets.

Recommended recalibration July 2025. Why should recalibration be carried out?

Recalibration of test equipment is a major requirement for quality management systems. All test weights vary with time due to wear and the collection of grime. The extreme of weight change varies with the environment the weights are used in, consequently periodic recalibration at regular intervals is required.

Signed



Weights

Unit A5, Greenway Business Park
Winslow Road, Great Horwood
Buckinghamshire, MK17 0NP

Tel: 01296 712171
sales@weights.uk.com
www.weights.uk.com

OIML +/- maximum permissible tolerance on calibration weights

Nominal Value	Class E ₁ mg	Class E ₂ mg	Class F ₁ mg	Class F ₂ mg	Class M ₁ mg	Class M ₂ mg
50kg	25	80	250	800	2,500	8,000
20kg	10	30	100	300	1,000	3,000
10kg	5.0	16	50	160	500	1,600
5kg	2.5	8.0	25	80	250	800
2kg	1.0	3.0	10	30	100	300
1kg	0.5	1.6	5.0	16	50	160
500g	0.25	0.8	2.5	8.0	25	80
200g	0.10	0.3	1.0	3.0	10	30
100g	0.05	0.16	0.5	1.6	5.0	16
50g	0.03	0.10	0.3	1.0	3.0	10
20g	0.025	0.08	0.25	0.8	2.5	8.0
10g	0.020	0.06	0.20	0.6	2.0	6.0
5g	0.016	0.05	0.16	0.5	1.6	5.0
2g	0.012	0.04	0.12	0.4	1.2	4.0
1g	0.010	0.03	0.10	0.3	1.0	3.0
500mg	0.008	0.025	0.08	0.25	0.8	2.5
200mg	0.006	0.020	0.06	0.20	0.6	2.0
100mg	0.005	0.016	0.05	0.16	0.5	1.6
50mg	0.004	0.012	0.04	0.12	0.4	-
20mg	0.003	0.010	0.03	0.10	0.3	-
10mg	0.003	0.008	0.025	0.08	0.25	-
5mg	0.003	0.006	0.020	0.06	0.20	-
2mg	0.003	0.006	0.020	0.06	0.20	-
1mg	0.003	0.006	0.020	0.06	0.20	-

Note: Non standard weights not listed, the next highest weight tolerance applies.

Weights

Directors: Simon Fenton, Lesley Fenton, Karen Fenton
VAT Number: GB 133 2359 02
Registered in England & Wales, Company 4787421
Registered Office: Investment House, 24 Vicarage Road, Winslow, Bucks MK18 3BE