

CERTIFICATE OF CALIBRATION

Issued by: Compli Group Ltd.

Unit 1, Cabot Business Village, Holyrood Close, Poole, Dorset, BH17 7BA
www.compli.uk

Telephone: 01202 656165

Email: sales@compli.uk

Date of Issue: 19 December 2024

Certificate Number: 414458

BS:EN ISO9001:2015

LRQA

REGISTERED COMPANY

LRQ 00005310

Customer:
Total Laboratory Services (TLS)
Unit 14C
Sunrise Business Park
Blandford Forum
Dorset
DT11 8ST

Page 1 of 1 Page
Approved Signatory
A. Garnett



Equipment: Inclinometer Fisco En 17,

Serial No: 96021

Conditions of Calibration: <70 %rh & 20 °C ±2 °C

Date of Calibration: 19 December 2024 **Procedure No.** Pro 0050

Acceptance Criteria: Manufacturer's Specification (±0.1 °)

Method of Calibration: Calibrated using angle blocks on a surface table

Results: As found unless specified

Nominal Angle	Inclinometer Reading Incline	Inclinometer Reading Decline	Tolerance ±	Result
0.0 °	0.0	0.0	0.2 °	Pass
1.0 °	1.0	0.9	0.2 °	Pass
3.0 °	3.0	2.9	0.2 °	Pass
9.0 °	8.9	8.9	0.2 °	Pass
27.0 °	26.9	26.8	0.2 °	Pass
41.0 °	40.8	40.8	0.2 °	Pass
90.0 °	89.8	-	0.2 °	Pass

Measurement Uncertainty: ±0.1 °

Measured results are traceable to national standards via equipment
PICL271, Certificate: 160499 (Cal due: 21/06/2027)

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%