

CERTIFICATE OF CALIBRATION

Issued by: **Compli Group Ltd.**

Unit 1, Cabot Business Village, Holyrood Close, Poole, Dorset, BH17 7BA
www.compli.uk

Telephone: 01202 656165

Email: sales@compli.uk

Date of Issue: 20 December 2024

Certificate Number: 414460

BS:EN ISO9001:2015

L R Q A

REGISTERED COMPANY

LRQ 00005310

Customer:
Total Laboratory Services (TLS)
Unit 14C
Sunrise Business Park
Blandford Forum
Dorset
DT11 8ST

Page 1 of 1 Page
Approved Signatory
A. Garnett



Equipment: Feeler Gauge Set Metric,

Serial No: 2693856/324

Conditions of Calibration: <70 %rh & 20 °C ±2 °C

Date of Calibration: 20 December 2024 **Procedure No.** Pro 0065

Acceptance Criteria: BS 957:2008

Method of Calibration: Calibrated using a horizontal metroscope

Results: As found unless specified. Feeler gauges rusted.

Feeler Indication mm	Measured Deviation 1=0.001 mm	Tolerance ± 1=0.001 mm	Result
0.05	+1.4	5.0	Pass
0.10	+1.4	5.0	Pass
0.15	+1.8	5.0	Pass
0.20	+1.8	5.0	Pass
0.25	+3.8	5.0	Pass
0.30	+0.4	8.0	Pass
0.40	+1.0	8.0	Pass
0.50	+2.2	8.0	Pass
0.60	+5.6	8.0	Pass
0.70	+6.8	10.0	Pass
0.80	+10.0	10.0	Pass
0.90	+2.0	10.0	Pass
1.00	+4.2	10.0	Pass

Measurement Uncertainty: ±0.002 mm

Measured results are traceable to national standards via equipment
PICL077, Certificate: 399518 (Cal due: 05/05/2026)

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%