




Total Laboratory Services Ltd

Inspection Certificate

Distance Plates Z95013

Number: 15-01-2025

Thickness measurement	Nominal	Average	Actual	
			Xmin	Xmax
Distance plate 0.6mm	0.6 ±0.05	0.58	0.56	0.59
Distance plate 1.0mm	1.0 ±0.05	1.00	0.98	1.02
Distance plate 1.5mm	1.5 ±0.05	1.50	1.49	1.51



* Certificate shows average for the plates in each category.


Material: Stainless Steel

Calibration Equipment:
Mitutoyo 150 mm Digital Calliper. Serial Number: 019160

Calibration Date: 18 March 2024 Re-Calibration Due: 18 March 2025

Certificate Number: 1854907

Date of Inspection: 15 January 2025
Blandford Forum, Dorset



Certified by: *P. Netherway*
Paul Netherway, Managing Director

Total Laboratory Services Ltd
14 C, Sunrise Business Park
Blandford Forum, Dorset DT11 8ST